Disintegration Tester

MODEL DIT-200

■ Feature

Model DIT-200, DIT-400
Available for 2 or 4 Baskets.
Stroke height is fixed at 55 mm and the number of strokes set at 30/min.
The instruments conforms to all pharma-copoeias a norms.
Heating system with control by the keypad and LED display; control accuracy ± 0.2°C.
The maximum disintegration time can be pre-set and there is a digital display for the elapsed test time.
The disintegration tester including full IQ and OQ documentation.

■ Specification

Basket	: 2 ea (as per K.P) With cover
Motor	: 25W (60:1) 1 ea / 2 ea
Time Monitor	: Digital 99H 59M 59s 1 ea / 2 ea Elapsed time (with alarm)
Water Bath	: SUS#304 Plate & safty glass 1 set Silicone calking & E.P. finished Heater AC220V 600W 1 ea
Temp. Controller	: Digital automatic PID control 1 set 0.1℃ display & set
Safety Deviced	: E.L.B S/W, Heater burn out Protector
Glass Thermometer	: 0 - 50℃ 3 ea
Spare Part	: Glass Tube (K.P.) 12 ea
Power	: AC220V 60HZ 600W

LABFINE
랩화인(주) 경기도 안양시 만안구 안양동 196 (유천팩토피아 218호)
Tel. 031-467-2525/6, Fax. 031–467–2527, E-mail. labfine@labfine.com
Homepage. http://www.labfine.com