Suppository Disintegration Tester

MODEL SDT-310

■ Feature

Model SDT-310

-Available for 3 test stations.
-Automatically turned 180° every 10 minutes.
-Membrane function keys for pre-selection of the test run-time.
-Heating system with control by the keypad and LED display, control accuracy ± 0.2℃.
-The SDT-310 is a basic suppository disintegration tester and complies with the latest EP specifications.
-The SDT-310 comes with each test station located inside a 4 liter glass beaker which are placed inside thermostatically heated water bath.
-Optional separately controllable magnetic stirrers can be placed in each test station.
- Power : AC220V 60HZ 1100W

■ Specification

Drive Unit : 3 set
Synchronous belt system
Basket : 3 ea
Beaker : 4 liter 3 ea
Temp. circulator : Digital auto PID 0.1℃ display
Circulation pump 100W
Heater AC220V 1KW
Bath : SUS#304 plate (with Safety Glass)
Time Monitor : Digital 0 - 120 min. with alarm
Motor : Automatically turned through 180 degrees every 10minutes
AC220V Geared Motor x 1 set
Optional : Magnetic Stirrer (with stirring bar) x 3 set

LABFINE
랩화인(주) 경기도 안양시 만안구 안양동 196 (유천팩토피아 218호)
Tel. 031-467-2525/6, Fax. 031–467–2527, E-mail. labfine@labfine.com
Homepage. http://www.labfine.com